



EU Declaration of Conformity – Pulsar Measurement PTFM 6.1 Portable Transit-Time Flow Meter

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Relevant Directive(s)	2014/30/EU – EMC directive and its amending directives 2014/35/EU – Low Voltage directive and its amending directives 2011/65/EU – RoHS directive and its amending directives
Manufacturer’s name	Greyline Instruments Canada Inc.
Manufacturer’s address	16456 Sixsmith Drive Long Sault, ON K0C 1P0 Canada
Apparatus	Transit-time flow meter including handheld electronics, transducer cables, and transducers
Models	PTFM 6.1 with SE16A, SE16B, and/or SE16C transducers
Type of equipment	Measurement and process control
Standards applied	EN 61326-1:2013 EMC, equipment class industrial EN 61010-1:2010+A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use European CISPR 11:2009 + A1:2010 / EN 55011:2009 + A1:2010, Class A, Group 1 FCC Part 15, Subpart B – Class A Unintentional Radiators for Use in Commercial and Industrial Areas

I declare that the apparatus named above has been tested and complies with the relevant sections of the above referenced standards & directives.

Signed:

Date: December 15, 2022
Rev. 1.0

Name & function

Ken Elander, Sr Product
Manager

Pulsar Measurement

