



## EU Declaration of Conformity – Pulsar Measurement MantaRay Portable Area Velocity Flow Meter

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Relevant Directive(s)	2014/30/EU – EMC directive and its amending directives 2014/35/EU – Low Voltage directive and its amending directives 2011/65/EU – RoHS directive and its amending directives
Manufacturer's name	Greyline Instruments Canada Inc.
Manufacturer's address	16456 Sixsmith Drive Long Sault, ON K0C 1P0 Canada
Apparatus	MantaRay flow meter including portable electronics, transducer with built-in cable, and break-out box
Models	MantaRay
Type of equipment	Measurement and process control
Standards applied	EN 61326-1:2013 EMC, equipment class industrial European CISPR 11:2009 + A1:2010 / EN 55011:2009 + A1:2010, Class A, Group 1

I declare that the apparatus named above has been tested and complies with the relevant sections of the above referenced standards & directives.

Signed:

A handwritten signature in black ink, appearing to read "Ken Elander".

Date: November 16, 2022  
Rev. 1.0

Name & function

Ken Elander, Sr Product  
Manager

Pulsar Measurement